

**WAC 16-102-030 Permanent records.** (1) A bound record book shall be provided for all permanent records.

(2) All calibration results and check tests for initial calibration or subsequent calibration shall be recorded in permanent record book.

(3) All tests for checking accuracy of calibration shall be recorded in permanent record book.

[Order 1133, § 16-102-030, filed 3/16/70.]