

WAC 246-227-130 Special requirements for enclosed radiography.

(1) Shielded room radiography systems and cabinet systems shall:

(a) Comply with all applicable requirements of this chapter and WAC 246-221-060;

(b) Be interlocked such that the exposure will terminate if a door or port accessible to individuals is opened during the exposure, except for those systems employing conveyor belts or sample ports; and

(c) Be tested for the proper operation of interlocks, high radiation area control devices or alarm systems, where applicable, at the beginning of each day of use. The results of these tests shall be recorded and maintained for three years.

(2) The registrant shall perform an evaluation, at intervals not to exceed one year, to determine conformance with this chapter and WAC 246-221-060. Records of each evaluation shall be maintained for three years.

[Statutory Authority: RCW 70.98.050. WSR 94-01-073, § 246-227-130, filed 12/9/93, effective 1/9/94.]