

WAC 246-240-587 Records of molybdenum-99, strontium-82, and strontium-85 concentrations. A licensee shall maintain a record of the molybdenum-99, strontium-82, or strontium-85 concentration tests required by WAC 246-240-160(2) for three years.

(1) The record must include, for each measured elution of technetium-99m, the ratio of the measures expressed as kilobecquerels of molybdenum-99 per megabecquerel of technetium-99m (or microcuries of molybdenum per millicurie of technetium), the time and date of the measurement, and the name of the individual who made the measurement.

(2) For each measured elution of rubidium-82, the ratio of the measures expressed as kilobecquerels of strontium-82 per megabecquerel of rubidium-82 (or microcuries of strontium-82 per millicurie of rubidium), or kilobecquerels of strontium-85 per megabecquerel of rubidium-82 (or microcuries of strontium-85 per millicurie of rubidium), the time and date of the measurement, and the name of the individual who made the measurement.

[Statutory Authority: RCW 70.98.050. WSR 13-11-021, § 246-240-587, filed 5/7/13, effective 6/7/13. Statutory Authority: RCW 70.98.050 and 70.98.080. WSR 09-06-003, § 246-240-587, filed 2/18/09, effective 3/21/09. Statutory Authority: RCW 70.98.050. WSR 06-05-019, § 246-240-587, filed 2/6/06, effective 3/9/06.]